

Calibration Report for Amplitude Control Linearity

A Nondestructive testing (NDT) calibration report format suitable for Ultrasonic Equipment calibrations. This is a sample report and may be required to be modified as per specific requirements.

NDT Procedure No: TNE-DOC-UT-03 Rev '0'



PETRO-TRACKING SERVICES LTD.®

An ISO 9001:2008 Certified Company

12 Irirhi Road, Benincity, Edo State, Nigeria | West Africa

Mobile: +2348108927239, +2348156210344

E-mail: info@petrotrackingserviceltd.com.ng, Website: www.petrotrackingserviceltd.com.ng

This UT calibration report & other NDT reports are available for free download at
www.petrotrackingserviceltd.com.ng




PETRO-TRACKING SERVICES LTD.®
An ISO 9001:2008 Certified Company

AMPLITUDE CONTROL LINEARITY

This procedure & other NDT procedures are available for free download at www.petrotrackingserviceltd.com.ng

Instrument Model	:		SL.No:
Manufacturer	:		
Applicable Code	:		

Sketch:



Indication Set at % Of Full Screen	dB Control Change	Accepted % Of Full Screen	Indication % Of Full Screen
80 %	-6 dB	32 to 48 %	
80 %	-12 dB	16 to 24 %	
40 %	+6 dB	64 to 96 %	
20 %	+12 dB	64 to 96 %	

	Tested By	Approved By	Third Party(if required)
Name			
Sign.			
Level			
Date			



E & P Solutions

NDT Training | Testing | Consultation

An ISO 9001 : 2008 Certified Company